

Instructions For use

Doc No.: RP-RA-03 RE02

This set of instructions for use was written with respect to the following non-invasive medical devices:

- Virox 70% Isopropyl Alcohol Wipes 1s (50 x 1s) (Product code: VR3926)
- Virox 70% Isopropyl Alcohol Wipes 1s (100 x 1s) (Product code: VR3927 and VR3927NKF)

The Basic UDI-DI of these non-invasive medical devices is 88863897ViroxV5.

Our intended use of these products is to kill 99.99% of germs. It is designed for disinfection and cleaning of hard surfaces. Not for end-point sterilization. Our indications of use are for disinfection of non-invasive medical devices. Not for end-point sterilisation. These products are designed for disinfection of all surfaces, such as metal, rubber and fabrics, and are meant to be disposed after single use. Avoid acrylic glass, acrylic surfaces and other plastic surfaces when using alcohol-based wipes.

The device is a medical device.

This product is for external use only. It should be stored in a cool place, away from naked flame and direct sunlight. Do not flush. The wipes can be used on general non-invasive medical devices as depicted in Figure 1, such as hospital beds, wheelchairs and thermometers. Make sure that all visible soiling on target surface is removed before disinfection. Wipe surface thoroughly with the impregnated wipe, ensure complete wetting and allow the surface to dry completely. Duration of using wet wipes is dependent on the drying time. Ensure that the antimicrobial efficiency of the product matches its intended use. Not for end-point disinfection of semicritical and critical medical devices.



Figure 1: Examples of non-invasive medical devices that are suitable for wet wipes



Doc No.: RP-RA-03 RE02

As a precaution, users should keep this product out of reach of children. This product is flammable, hence keep away from heat, sparks and open flame. Avoid contact with eyes. Store at room temperature. Wearing of gloves (eg. Nitrile rubber) is recommended. First aid: if in contact with eyes; hold eyes open and rinse slowly and gently with water for 15-20 minutes. If contact lenses are present, remove them immediately and rinse eyes. Call a poison control center for treatment advice. Any serious incident that has occurred in relation to the device should be reported to manufacturer and local authority of the country.



Doc No.: RP-RA-03 RE02

DO NOT PRINT THIS PAGE

Author	Approval
	R&D/ Regulatory Manager
	Trans
Jogi Soh	Chey Kok Leong

Revision	History of changes to document	Released date
00	Original	14-09-2020
01	Change in intended use	23-09-2020
02	Add in the item listed in Annex I, Chapter III 23.4 (z): with additional caution and 23.2 (q): indication that the device is a medical device	11-04-2022